RETURN TO TECHNICAL SERVICE BULLETIN TABLE OF CONTENTS
RETURN TO CD-ROM TABLE OF CONTENTS

Fabius GS
TECHNICAL SERVICE BULLETIN # 3

Product:	Fabius GS
Date:	October 25, 2002
Re:	Software 1.32
Reference Doc:	Product Bulletin # 136; All Fabius GS field units with software 1.31 (attached)
Reason/Rational:	Desflurane Compensation Correction
Solution/ Procedure:	Perform Software 1.32 download on all current Fabius GS field units containing software 1.31.
Devices affected:	All field units shipped with or upgraded to software 1.31. See attached. Domestic units with 1.20 and export units containing 1.2x are not affected.
Timing/When:	At next service or Periodic Manufacturer's Service.
Additional Info:	Software downloads from 1.2x to 1.32 shall only be performed using upgrade kit P/N 4117917. Refer to PB #136 for details.
Costs:	Software 1.32 is available to Fabius GS customers with software 1.31. at no charge. Labor – Warranty: With reference of this TSB.
Feedback:	Complete Return Form (attached) and fax to DrägerService at (215) 721-7584 upon software completion/account.
Ordering Info/Parts	Software 1.32 can be obtained through the NAD Machine Software Database or; Contact Technical Support at techsupport@draegermed.com.
Distribution:	DrägerService & All Authorized Service Organizations.

If you have any questions, please contact Technical Support at: Ph: 1-800-543-5047, Fax: 1-215-721-5789 or techsupport@draegermed.com

DrägerService
3122 Commerce Drive
Telford, PA 18969
Tel: (215) 721-5402
(800) 543-5047
Fax: (215) 721-5784
Web: www.draegermed.com
DrägerService is a division of
Draeger Medical

RETURN TO TECHNICAL SERVICE BULLETIN TABLE OF CONTENTS
RETURN TO CD-ROM TABLE OF CONTENTS

Return Form - Software 1.32 Update

Distributor: _____ Customer: _____

Country: _____ Address: _____

City: _____

Machine S/N	Machine S/N	Machine S/N
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____

Fax To:
DrägerService
Attn: Al Myers
(215) 721-5784

Date of Update _____
Signature (TSR) _____
Name (Print) _____

DrägerService
3122 Commerce Drive
Telford, PA 18969
Tel: (215) 721-5402
(800) 543-5047
Fax: (215) 721-5784
Web: www.draegermed.com
DrägerService is a division of
Draeger Medical

RETURN TO TECHNICAL SERVICE BULLETIN TABLE OF CONTENTS
RETURN TO CD-ROM TABLE OF CONTENTS

DrägerService
3122 Commerce Drive
Telford, PA 18969
Tel: (215) 721-5402
(800) 543-5047
Fax: (215) 721-5784
Web: www.draegermed.com
DrägerService is a division of
Draeger Medical